Time-Series at the Edge: Tiny Separable CNNs for Wearable Gait Detection and Optimal Sensor Placement

Andrea Procopio*1,4, Marco Esposito², Sara Raggiunto², Andrey Gizdov^{1,3}, Alberto Belli², Paola Pierleoni²

¹Harvard University, Cambridge, MA, USA ²Polytechnic University of Marche (UNIVPM), Italy ³Weizmann Institute of Science, Rehovot, Israel ⁴Bocconi University, Italy {aprocopio, andreygizdov}@fas.harvard.edu, {m.esposito, s.raggiunto, a.belli, p.pierleoni}@staff.univpm.it

Abstract

We study on-device time-series analysis for gait detection in Parkinson's disease (PD) from short windows of triaxial acceleration, targeting resource-constrained wearables and edge nodes. We compare magnitude thresholding to three 1D CNNs for time-series analysis: a literature baseline (separable convolutions) and two ultra-light models—one purely separable and one with residual connections. Using the BioStampRC21 dataset, 2 s windows at 30 Hz, and subjectindependent leave-one-subject-out (LOSO) validation on 16 PwPD with chest-worn IMUs, our residual separable model (Model 2, 533 params) attains PR-AUC = 94.5%, F_1 = 91.2%, MCC = 89.4%, matching or surpassing the baseline $(5,552 \text{ params}; PR-AUC = 93.7\%, F_1 = 90.5\%, MCC =$ 88.5%) with $\sim 10 \times$ fewer parameters. The smallest model (Model 1, 305 params) reaches PR-AUC = 94.0%, F_1 = 91.0%, MCC = 89.1%. Thresholding obtains high recall (89.0%) but low precision (76.5%), yielding many false positives and high inter-subject variance. Sensor-position analysis (train-on-all) shows chest and thighs are most reliable; forearms degrade precision/recall due to non-gait arm motion; naive fusion of all sites does not outperform the best single site. Both compact CNNs execute within tight memory/latency budgets on STM32-class MCUs (sub-10 ms on lowpower boards), enabling on-sensor gating of transmission/storage. Overall, ultra-light separable CNNs provide a superior accuracy-efficiency-generalization trade-off to fixed thresholds for wearable PD gait detection and underscore the value of tailored time-series models for edge deployment.

Introduction

Parkinson's Disease (PD) is now one of the fastest-growing neuro-degenerative disorders in the world (Chaudhuri et al. 2024). Despite optimal treatment, the progression of the disease leads to a gradual deterioration in motor and cognitive disorders, with inevitable compromise in patients' quality of life. Several works in the literature study gait disabilities using wearable, small, and non-invasive sensors. To design personalized interventions and overcome the limitations related to self-reported questionnaires and diaries, the focus is to improve clinical and home care through the additional information on daily life performance

provided by Inertial Measurement Units (IMUs) (Corrà et al. 2021). It is essential to develop algorithms for rapid and precise detection of motor activity in subjects with Parkinson's Disease to intervene promptly in the detection of anomalies to safeguard the subject's health. To carry out continuous and constant monitoring of PD subjects, particularly in a domestic and unsupervised environment, the first fundamental step is the detection of the activities carried out by the subject (Muthukrishnan, Abbas, and Krishnamurthi 2020; Corrà et al. 2021). In particular, gait activities are of interest due to the fact that PD severely affects the gait cycle and, therefore, gait cycle impairments should be monitored and detected promptly. Following gait detection, it is advisable to perform gait analysis to detect all possible anomalies such as fluctuations, dyskinesias, etc (Sigcha et al. 2020; Mancini et al. 2021). The main studies either process the signal offline or use Cloud-based services. Cloud-based architecture needs data to be moved to the data centers, and are typically employed to support data long-term analyses. However, this scenario is contrary to IoT application requirements, such as limit costs, memory footprint, processing, and communication resources (Kianoush et al. 2023). In recent years, there has been a shift towards edge computing, which means processing data on devices closer to the data sources, rather than relying on cloud-based services to reduce the amount of data to be sent to the cloud by the IoT devices (Fathalla et al. 2022). This shift has also been enabled by advances in embedded systems, as well as tools and accelerators that make it possible to run machine learning and artificial intelligence algorithms on low-power devices (Dini et al. 2024).

In this paper, an in-depth investigation of gait recognition techniques is conducted as a first step towards developing a decentralized architecture for Parkinson's disease monitoring where distributed processing happens across wearable devices, Edge nodes, and the Cloud. The proposed gait recognition algorithm aims to reduce data transmission through local data processing, optimizing the system's energy efficiency and improving the quality of stored data by eliminating redundant or irrelevant data samples. Compared to previous works, we provide a comprehensive evaluation that compares simple threshold-based methods to separable CNNs based frameworks, including a Residual

^{*}Corresponding author: aprocopio@fas.harvard.edu Copyright © 2026, Association for the Advancement of Artificial Intelligence (www.aaai.org). All rights reserved.

Network structure. Our evaluation covers multiple sensor positions and combinations, assesses the methods' performance across different patients (Leave-One-Subject-Out validation), and examines computational complexity and inference time on a target MCU.

In this work, we make the following contributions:

- We introduce two lightweight configurations of 1D-CNNs based architectures achieving a better complexityperformance tradeoff and generalization via F1-score compared to a reference, larger baseline. Through the LOSO evaluation, we show that is possible to design a super-light architecture that matches performances of models 18x times biggers while having extremely light inference times in tasks of binary gait detection on microcontrollers.
- 2. We discuss flaws of a threshold based approach, mainly concerning an inability for the framework to generalize among different subjects and sensor positions, and to maintain a stable accuracy level.
- 3. We conduct an in-depth investigation of models' performances and generalization abilities based on the position of data acquisition sensors. We show that data acquired from accelerometers placed in the chest area is consistently less noisy and easier to classify for all the frameworks evaluated, while thighs, and in particular forearms, appear to show greater inconsistency and variance.

Methods

Our goal is to investigate optimal architectures for gait recognition via efficient wearable sensors. Gait recognition can be utilized for contextual activation of efficient data transmission/storage to limit bandwidth and power consumption, which is especially useful for daily monitoring.

Data and pre-processing

Dataset description The raw data utilized in this study was obtained from the "BioStampRC21" dataset (hereafter, BioStamp) by (Adams et al. 2020), which is accompanied by clinical metadata including patient IDs and health statuses (Patients with Parkinson's Disease (PwPD) or Control). BioStamp gathers data from 17 PwPD and 17 control subjects of matching ages over an average time of 45.4 hours. The PwPD sample consists of 7 women and 10 men with a mean age of 66.4 years (11.3 sd), while the agematched sample consists of 13 women and 4 men with a mean age of 64.0 years (9.9 sd). Acceleration data in each individual subject is recorded from 5 different sensor position: chest (ch), left thigh (ll), right thigh (rl), left forearm (lh), right forearm (rh). A clinical annotation file for each subject provides a description of the clinical test being conducted at each timestamp during data collection.

Preprocessing Data annotations were categorized into "gait" and "non-gait" events based on the nature of the clinical test. For example, gait-related activities included UPDRS and UHDRS test versions of "Timed Up and Go", "10 Meter Walk Test", "Gait", and "Tandem Walking". Events that

did not comprise the selected gait activities were categorized as non-gait. The only annotation excluded from both criteria was the "UHDRS 15 - Retropulsion Pull Test", due to ambiguity in classifying the event and a negligible relevance both in the dataset and for the representation of domestic activities in PwPD. Sensor data were sourced from accelerometers placed at the five previously described body positions: ch, ll, rl, lh, rh. Each participant's sensor data underwent column standardization, renaming original columns to maintain consistency across the dataset. Subsequently, the accelerometer readings were uniformly resampled from 31.25 Hz to 30 Hz to allow the use of integer-length windows.

Windowing Data corresponding to gait and non-gait intervals were segmented using timestamps derived from clinical annotations. Each sensor recording was filtered according to these intervals to isolate relevant data segments. Windows of accelerometer data were created using a sliding-window approach with a window size of 60 samples (2 seconds) and a step size of 15 samples, resulting in 75% overlap uniformly across both classes. Each windowed data segment was standardized by subtracting the mean, ensuring zero-centered data for analysis. Finally, data segments were labeled according to the sensor position, the type of activity (gait or non-gait) and the participant's clinical status (PD or Control).

Final datasets Processed data from each sensor were compiled into structured dictionaries, organized by participant IDs and sensor positions. This led to the creation of 5 datasets, based on sensor positions. During the dataset generation process, subject "ID38" was excluded from the dataset generation process due to data quality issues. Acceleration data from all 5 sensors were aggregated across each participant and concatenated to form a comprehensive dataset. Each dataset (the 5 per-sensor datasets and the comprehensive dataset) was divided into training set (60%), a validation set (20%), and test set (20%) using stratified random splitting to maintain a consistent label distribution. The stratification ensured balanced representation of gait and non-gait labels in both training and validation datasets.

Leave-One-Subject-Out (**LOSO**) . To prepare the dataset for LOSO cross-validation, only PwPD were selected for the analysis, resulting in a final pool of 16 subjects: IDs = [6, 10, 12, 13, 15, 17, 23, 24, 25, 33, 35, 36, 40, 42, 44, 63]. For each of the resulting PwPD, data from that participant served as the test set, while the remaining participants data constituted training (70%) and validation (30%) dataset. Accelerometer data only from the chest sensor were extracted for this analysis as they proved to be the most stable from our sensor-dependent analysis. Every datasets obtained – LOSO, sensor-specific and combined – is composed by 60 x 3 shape windows, representing the 2 second windows with triaxial accelerometer data sampled at 30hz.

Gait detection algorithms

Magnitude thresholding. While AI methods offer replicability and generalization, simple thresholding has proven able to distinguish between gait and non-gait while still

Component	Configuration
Input	(60, 3)
SepConv1	filters: 100, kernel: 10, activation: ReLU
Pooling	MaxPool1D (pool size: 3): length $60 \rightarrow 20$
SepConv2	filters: 40, kernel: 10, activation: ReLU
Aggregation	GlobalAvgPooling1D, output dim: 40
Regularization	Dropout, $p = 0.5$
Classifier	Dense (2), softmax

Table 1: Reference baseline architecture.

maintaining extremely low computational cost (Borzì, Sigcha, and Olmo 2023; Soumma et al. 2024). As such, a comparison between the 1D-CNN models and a threshold-based detector is carried out. Thresholding is applied to the magnitude of the acceleration signal. For the j^{th} window, magnitude is computed as follows:

$$A_j = \sum_{i=1}^n \left(a_{x_i}^2 + a_{y_i}^2 + a_{z_i}^2 \right)$$

where n is the number of samples in the window, and a_{x_i} , a_{y_i} , and a_{z_i} are the measured acceleration components along the 3 measurement axes for the i^{th} time step. The identification of gait windows is performed by evaluating if the magnitude of the window is above the threshold. Consistently to previous works Borzì, Sigcha, and Olmo 2023, the optimal threshold is selected as the value that maximizes the F1 score across the training dataset. In the evaluation process, all windows with magnitude higher than the optimal threshold are classified as gait, and conversely the windows with a lower magnitude are classified as non-gait.

Deep Learning frameworks with CNNs.

In this section we describe and evaluate three different onedimensional Convolutional Neural Networks (1D-CNNs) architectures for Time Series Analysis.

Reference 1D-CNN Baseline Following Borzì, Sigcha, and Olmo (2023), who adapt the DeepFog CNN of Bikias et al. (2021) by replacing standard 1D convolutions with separable 1D convolutions to reduce complexity, we adopt the same two-block 1D SepConv CNN as our baseline, instantiated at our window length (60 samples) and with a two-class softmax head. Table 1 shows the architecture. Both SepConv layers include bias terms and use acausal padding before the ReLU activation.

Parameter profile: SepConv1: 430 params; SepConv2: 5,040; Dense: 82; total 5,552.

Model 1: Super-Light Separable Convolutions Network

The first architecture we propose is a highly compact variant of the baseline (retaining depthwise and pointwise separable 1D convolutions) but with a very reduced parameters count: fewer filters (8 and 16 instead of 100 and 40), smaller kernels (5 and 7 instead of 10), less aggressive pooling ($\times 2$ instead of $\times 3$), Batch Normalization (BN) after each convolution with biases disabled, and the same global average pooling + softmax head. Furthermore, we adopt an increasing

Component	Configuration
Input	(60, 3)
SepConv1	filters: 8, kernel: 5
Post-processing	$BN \rightarrow ReLU \rightarrow MaxPool1D$ (pool size: 2)
SepConv2	filters: 16, kernel: 7
Post-processing	$\mathrm{BN} o \mathrm{ReLU}$
Aggregation	GlobalAvgPooling1D, output dim: 16
Classifier	Dense (2), softmax
Initialization	He normal (depthwise and pointwise)

Table 2: Model 1 architecture.

Component	Configuration
Input	(60,3)
ResBlock1 (main)	SepConv1D: filters: 8, kernel: 9
Post-processing	$BN \rightarrow ReLU \rightarrow AvgPool1D$ (size: 2)
ResBlock1 (skip)	$1 \times 1 \text{ Conv} + \text{BN} (3 \rightarrow 8)$
Residual	$\mathrm{Add} ightarrow \mathrm{ReLU}$
ResBlock2 (main)	SepConv1D: filters: 16, kernel: 9
Post-processing	$BN \to ReLU$
ResBlock2 (skip)	$1 \times 1 \text{ Conv} + \text{BN } (8 \rightarrow 16)$
Residual	$\mathrm{Add} o \mathrm{ReLU}$
Aggregation	GlobalAvgPooling1D, output dim: 16
Classifier	Dense (2), softmax
Initialization	He normal (depthwise, pointwise, 1×1)

Table 3: Model 2 architecture.

number of channels per layer $(8\rightarrow16)$ rather than the baseline's decreasing one $(100\rightarrow40)$. This follows the common CNN practice of widening the network with depth to capture more complex features and mitigate information bottlenecks in later layers (Simonyan and Zisserman 2014; He et al. 2016; Sandler et al. 2018).

Parameter profile. Params (including BN γ , β): SepConv1: 39; BN1: 16; SepConv2: 184; BN2: 32; Dense: 34; total **305**.

Model 2: SuperLight Residual Separable CNN. The second architecture we analyse, Model 2, keeps the same lightweight design footprint as Model 1 (separable 1D convolutions and two stages) but changes two aspects: (i) each stage is a residual block with an identity skip (projected with a 1×1 convolution when channel dimensions differ), and (ii) both convolutions use a larger kernel (k=9). The first block downsamples with average pooling (factor 2); the second does not downsample. Each block has convolutions with bias disabled, followed by BatchNorm and ReLU; the residual sum is followed by a ReLU. A global average pooling layer and a two-class softmax head complete the model.

Parameter profile. With input (60,3) the trainable parameters are: Block 1 (SepConv+BN+projection) = 107, Block 2 = 392, Dense head = 34, totaling **533** parameters.

Training Protocol

All models share the same pipeline and training hyperparameters. Due to compatibility requirements all models have Softmax heads and use categorical cross-entropy and one-hot labels.

Data split (LOSO). We use Leave-One-Subject-Out (LOSO) evaluation. For each held-out subject s, we train on the remaining subjects, split a validation subset from the training portion (as provided by the dataset loader), and test on s. All choices below are applied identically in every fold and to every model.

Optimization. We train using AdamW, Loshchilov and Hutter (2019) – Adaptive Moment Estimator with Decoupled Weight Decay – learning rate 1×10^{-3} , weight decay 1×10^{-4} , $\beta_1 = 0.9$, $\beta_2 = 0.99$, $\epsilon = 10^{-8}$. Batch normalization and bias parameters are excluded from weight decay. While training is capped at 300 epochs with batch size 256, we monitor validation PR-AUC (Average Precision) and use Early Stopping with patience 12, restoring the best weights. We apply Reduce Learning Rate on Plateau on the same metric (factor 0.2, patience 7, minimum learning rate 1×10^{-5}). Furthermore, we apply class-balanced weights based on the *effective number of samples* (Cui et al. 2019). Let n_c be the number of training samples for class $c \in \{0,1\}$ and $\beta = 0.999$. The effective number is $E(n_c) = \frac{1-\beta^{n_c}}{1-\beta}$; the unnormalized weight is $w_c' = \frac{1}{E(n_c)} = \frac{1-\beta}{1-\beta^{n_c}}$. We normalize to unit average weight,

$$w_c = \frac{w_c'}{\frac{w_0' n_0 + w_1' n_1}{n_0 + n_1}},$$

and pass (w_0, w_1) during training for loss computation. After training in each fold, we compute positive-class scores on the validation set and select a single threshold τ^* that maximizes validation F1 by sweeping 401 uniformly spaced values in [0,1]. We then $fix\ \tau^*$ and evaluate the test subject with no further tuning. Note on reproducibility: Implementation uses TensorFlow–Keras (v2.15) and scikit-learn metrics for PR-AUC and F1. A fixed random seed (n=12) is set for all the training scripts. All training and testing settings are identical across models.

Deployment evaluation. After training, they were exported on the ST Edge AI Developer Cloud tool, which provides a framework for the evaluation of DL models for their deployment on various STM32 boards and MCUs. In particular, it provides estimations on memory occupancy and inference time for a given target MCU.

Evaluation Metrics

The dataset is imbalanced (about 80% non-gait, 20% gait). We therefore prioritize metrics that are informative under class imbalance and report both threshold-free and thresholded performance. All metrics are computed per LOSO fold; tables report the mean across folds (per-subject values in the supplement).

Notation. Let TP, TN, FP, FN denote the confusion counts for the positive class (gait), then we compute the following. PR-AUC (Average Precision),i.e., the area under the precision–recall curve (computed as Average Precision); threshold-free and robust to class imbalance. F1 at fixed threshold, i.e., the harmonic mean of precision and recall at the model's fixed decision threshold tau, selected on validation by maximizing F1:

$$F1 = \frac{2 PPV TPR}{PPV + TPR}, \quad PPV = \frac{TP}{TP + FP}$$

$$TPR = \frac{TP}{TP + FN}, \quad TNR = \frac{TN}{TN + FP}.$$

Furthermore, in the summary table 4 we include the Geometric Mean of TPR and TNR (GM), the chance-corrected agreement (Cohen's K), the Accuracy (kept for reference but de-emphasized under class imbalance), and the MCC Matthews correlation coefficient, which provides a single summary that captures whether a model balances both classes rather than succeeding by majority-class guessing.

$$Accuracy = \frac{TP + TN}{TP + TN + FP + FN}$$

$$MCC = \frac{TP \cdot TN - FP \cdot FN}{\sqrt{(TP + FP)(TP + FN)(TN + FP)(TN + FN)}}$$

Reporting. For each fold we fix the threshold τ^* on validation by maximizing F1, then evaluate the test subject with that τ^* . We report PR-AUC (threshold-free), and at τ^* we report F1, Precision (PPV), Recall (TPR), Specificity (TNR), MCC, and Accuracy.

Results

Subject-Independent LOSO on Chest IMU

Table 4 reports LOSO means over 16 PwPD using the chest IMU (2 s, 60 samples, 3 axes). Both proposed compact CNNs outperform the Baseline model while using 90–95% fewer parameters: $Model\ 2$ reaches F1=91.2%, PR–AUC=94.5%, MCC=89.4%; $Model\ 1$ attains F1=91.0%, PR–AUC=94.0%, MCC=89.1%. The Baseline architecture yields F1=90.5%, PR–AUC=93.7%, MCC=88.5%. Gains vs. Baseline are small but consistent across PPV (Model 2: 94.0 vs. 93.3) and TPR (Model 2: 89.1 vs. 88.3). MCC trends mirror F1/PR–AUC, indicating a better balance of false positives/negatives under the \approx 80/20 non-gait/gait skew.

The threshold baseline has good recall but low precision and unstable across subjects. Averaged over LOSO (Table 4), it achieves TPR=89.0% but only PPV=76.5%, for F1=81.5% and MCC=77.7%. Per-subject variability is high (F1 $\sigma=9.9$; Table 8), reflecting numerous false positives on some patients. Generalization across positions is also weak in the subject-mixed setting (Table 10): on forearms, PPV falls behind despite high recall (Left Arm: PPV=43.0%, F1=57.1%; Right Arm: PPV=42.2%, F1=57.0%), possible

Table 4: Average LOSO performance. Metrics are percentages; higher is better.

Metric	Model 1	Model 2	Baseline	Threshold
Accuracy	96.6	96.7	96.4	92.4
Specificity	98.5	98.6	98.5	93.6
Recall	88.9	89.1	88.3	89.0
Precision	93.7	94.0	93.3	76.5
F1-Score	91.0	91.2	90.5	81.5
MCC	89.1	89.4	88.5	77.7
GM	93.5	93.6	93.1	90.9
Cohen-K	88.9	89.2	88.3	76.8
PR-AUC	94.0	94.5	93.7	_

due to non-gait arm motions that can mimic the periodicity of motion. Despite having only 305 (Model 1) and 533 (Model 2) parameters compared to the Baseline's 5,552, both compact models match or surpass it on LOSO chest. Relative to Baseline, Model 2 gains +0.7 F1, +0.8 PR-AUC, and +0.9 MCC; Model 1 gains +0.5 F1, +0.3 PR-AUC, +0.6 MCC. Versus the threshold method, Model 2 improves PPV by +17.5 points (94.0 vs. 76.5) at similar recall (89.1 vs. 89.0), reducing false positives without relevant effects in sensitivity. Tables 5-7 detail per-subject results, also showing standard deviation (σ) and average values (μ) for all metrics. Difficult cases appear for all CNNs (e.g., IDs 25, 33, 6), where TPR is lowered while PPV remains high. Several subjects approach maximum values across all metrics (e.g., IDs 36, 40, 15, 13). Thresholding, although in general worse across metrics, achieves better results for the same subjects. However, it exhibits larger variance across metrics (F1 $\sigma = 9.9$)) with very unstable PPV, showing a higher susceptibility to false positives. While the simplest and most lightweight method, it is therefore not suitable for applications where misclassification cost is high (e.g., when it is needed to transmit or store gait windows).

Table 10 shows the result on the evaluation of different sensor positions, specifically after training the models on the entire dataset and then testing each position separately. Chest and thighs are consistently strongest across models. For the residual separable CNN (Model 2), left-thigh is best (F1 = 88.7%), followed by chest (F1 = 87.5%) and right-thigh (F1 = 84.7%). The lightweight separable CNN (Model 1) follows the same pattern (left-thigh 88.5%, chest 84.2%). The literature baseline also peaks on the left-thigh (89.1%) and chest (88.4%). Forearms underperform for all CNNs (Model 2: 76.0-81.8% F1), and the threshold baseline shows the highest precision drop on forearms (PPV = 43.0%/42.2%; F1 = 57.1%/57.0% for left and right, respectively) despite high recall. The Threshold method performs significantly worse when tested on the fused 'All' sensor data (F1 = 67.7%) compared to the CNN models (Model 2 F1 = 83.8%), indicating its poor generalization across sensor sites.

Compute efficiency and deployability

We profile three STM32 targets (L4R9I-DISCO, F401RE, H7S78-DK). Table 9 reports memory, MACs, and per-

Table 5: LOSO evaluation for Model 1 (lightweight SepCNN). Metrics are percentages. μ and σ denote mean and standard deviation across subjects.

ID	Acc (%)	TNR (%)	TPR (%)	PPV (%)	F1 (%)
6	92.8	97.8	78.2	92.7	84.8
10	98.0	100.0	90.8	100.0	95.2
12	95.0	96.8	86.6	84.6	85.6
13	98.5	99.8	92.4	98.8	95.5
15	96.8	98.3	91.8	94.2	93.0
17	97.9	98.9	93.4	95.1	94.2
23	98.3	99.6	91.7	97.8	94.7
24	98.2	99.2	89.3	92.6	90.9
25	93.4	97.6	78.1	90.2	83.7
33	95.2	99.6	77.2	97.9	86.4
35	99.1	99.7	95.8	98.3	97.0
36	99.9	100	99.0	100.0	99.5
40	96.1	98.4	84.1	91.2	87.5
42	95.2	98.7	84.1	95.4	89.4
44	94.4	94.5	93.8	79.1	85.8
63	97.2	97.6	95.4	91.3	93.3
$\overline{\mu}$	96.6	98.5	88.9	93.7	91.0
σ	2.1	1.4	6.8	5.7	4.9

Table 6: LOSO evaluation for Model 2 (ResNet). Metrics are percentages. μ and σ denote mean and standard deviation across subjects.

ID	Acc (%)	TNR (%)	TPR (%)	PPV (%)	F1 (%)
6	92.1	98.0	75.1	93.1	83.2
10	97.9	99.9	90.8	99.6	95.0
12	96.2	98.1	86.9	90.5	88.6
13	98.7	99.5	94.7	97.6	96.1
15	98.1	98.9	95.7	96.2	95.9
17	98.1	99.1	94.1	95.9	95.0
23	97.8	99.3	90.3	96.3	93.2
24	98.9	99.3	95.2	94.1	94.7
25	91.3	97.9	67.6	90.2	77.3
33	95.7	99.7	79.3	98.5	87.8
35	99.2	99.8	95.8	99.1	97.4
36	99.6	99.5	100.0	97.1	98.5
40	96.9	98.3	89.8	90.9	90.4
42	95.8	99.3	84.6	97.5	90.6
44	93.5	94.3	89.9	77.6	83.3
63	96.7	97.1	95.2	89.5	92.3
$\overline{\mu}$	96.7	98.6	89.1	94.0	91.2
σ	2.5	1.4	8.6	5.5	5.9

window latency. The Baseline is largest and slowest (133,755 MACs; 13.75 KB RAM; H7: 3.22 ms; L4/F4: 26.68/32.72 ms). Model 1 is the most compact (9,879 MACs; 4.87 KB RAM) and fastest (H7: 0.45 ms; L4/F4: 3.42/4.18 ms). Model 2 offers the best accuracy–efficiency balance (18,303 MACs; H7: 0.70 ms; L4/F4: 5.50/6.80 ms), delivering slightly higher F_1 than Baseline with $\sim 7.3 \times$ fewer MACs. All models fit without external flash/RAM, enabling on-sensor gating. Because dynamic energy on MCUs scales roughly with operation count, lower MACs imply lower energy per inference (Li et al. 2025). By contrast,

Table 7: LOSO evaluation for the Baseline model. Metrics are percentages. μ and σ denote mean and standard deviation across subjects.

ID	Acc (%)	TNR (%)	TPR (%)	PPV (%)	F1 (%)
10	91.2	98.0	71.8	92.8	80.9
12	97.8	100.0	89.7	100.0	94.6
13	94.4	96.6	83.7	83.5	83.6
15	98.7	99.8	92.8	99.2	95.9
17	97.0	97.4	95.4	91.6	93.5
23	98.3	99.1	95.2	95.9	95.6
24	97.9	99.4	90.3	97.0	93.6
25	98.3	98.9	92.9	90.7	91.8
33	92.5	97.1	76.1	87.9	81.6
35	95.4	99.3	79.3	96.5	87.1
36	99.4	99.8	96.6	99.1	97.9
40	99.9	100	99.0	100	99.5
42	96.1	98.8	82.1	93.1	87.3
44	95.0	99.0	82.7	96.2	89.0
6	94.4	95.4	89.9	81.1	85.3
63	96.3	96.5	95.2	87.7	91.3
$\overline{\mu}$	96.4	98.5	88.3	93.3	90.5
σ	2.5	1.4	8.0	5.9	5.8

Table 8: LOSO evaluation for the Threshold baseline. Metrics are percentages. μ and σ denote mean and standard deviation across subjects.

ID	Acc (%)	TNR (%)	TPR (%)	PPV (%)	F1 (%)
6	83.4	93.4	55.0	74.3	63.2
10	94.2	95.2	90.8	83.8	87.2
12	91.2	91.5	89.7	68.3	77.5
13	95.2	95.0	96.2	79.3	86.9
15	96.7	95.8	99.7	87.6	93.3
17	93.9	92.8	98.9	76.0	85.9
23	96.4	98.2	87.6	90.7	89.1
24	97.5	97.2	100	80.0	88.9
25	84.5	93.5	52.0	69.1	59.3
33	90.3	92.5	81.3	72.5	76.6
35	92.9	92.2	96.6	69.1	80.6
36	96.6	96.0	100	81.1	89.6
40	94.1	94.3	93.1	76.1	83.7
42	92.9	93.9	89.7	82.4	85.9
44	84.6	81.9	96.9	54.1	69.4
63	94.0	93.6	95.7	79.4	86.8
$\overline{\mu}$	92.4	93.6	89.0	76.5	81.5
σ	4.5	3.6	14.8	8.8	9.9

fixed-threshold detection has cost that grows linearly with input length and is effectively negligible on our windows (Borzì, Sigcha, and Olmo 2023), but its poor precision and high inter-subject variance (Table 8) can make downstream false-positive costs (radio, storage, block activation) dominate.

Conclusion

This work evaluated 1D convolutional neural network architectures and a threshold baseline for gait-event detection

Table 9: Model Metrics and Inference Latency

Model	Flash	RAM	MACs	Inference Latency (ms)			
Model	(KB)	(KB)	(M)	L4R9I-DISCO	F401RE	H7S78-DK	
Model 1	20.62	4.87	9879	3.42	4.18	0.45	
Model 2	23.86	11.93	18303	5.50	6.80	0.70	
Baseline	42.12	13.75	133755	26.68	32.72	3.22	

Table 10: **Train-on-all, per-sensor evaluation (subject-mixed).** Metrics are percentages. Acc = accuracy; TNR = specificity; TPR = recall; PPV = precision.

Model	Sensor	Acc	TNR	TPR	PPV	F1
	Chest	94.5	99.5	74.2	97.3	84.2
	Left arm	91.4	96.1	72.0	81.8	76.6
Model 1	Right arm	92.4	96.3	76.7	83.5	80.0
Model 1	Left leg	95.8	99.0	82.6	95.3	88.5
	Right leg	94.1	98.6	75.6	93.2	83.4
	All	93.7	97.9	76.3	90.0	82.5
	Chest	95.5	99.1	80.5	95.8	87.5
	Left arm	91.3	96.3	70.7	82.3	76.0
Model 2	Right arm	93.2	97.1	77.3	86.9	81.8
Model 2	Left leg	95.8	98.7	83.8	94.2	88.7
	Right leg	94.5	98.6	77.6	93.1	84.7
	All	94.1	98.0	78.1	90.5	83.8
	Chest	95.8	99.1	81.9	95.9	88.4
	Left arm	92.0	96.9	71.9	84.8	77.8
Baseline	Right arm	93.5	97.8	76.2	89.4	82.3
Daseille	Left leg	96.0	98.9	83.8	95.1	89.1
	Right leg	94.9	98.8	78.7	94.2	85.7
	All	94.4	98.3	78.6	92.0	84.8
	Chest	91.7	95.0	78.3	79.2	78.8
	Left arm	75.1	72.7	84.9	43.0	57.1
Threshold	Right arm	73.9	70.6	87.6	42.2	57.0
THESHOLD	Left leg	90.1	90.0	90.7	68.9	78.3
	Right leg	88.2	88.3	87.9	64.8	74.6
	All	83.9	83.4	85.9	55.8	67.7

in people with Parkinson's disease using accelerometer data from different body locations. It was shown how chest data clearly outperformed other sensor positions, which included arms and legs. The goal is to detect gait windows on-sensor to replace raw data streams to improve bandwidth, storage, and energy. In leave-one-subject-out validation, all CNNs outperformed the threshold baseline, which in general shows poor generalization. A detailed deployment analysis for embedded systems demonstrated that all models run within memory and latency constraints of representative STM32 microcontrollers. Model 2, which adds residual connections on top of the separable-convolution design of Model 1 and the baseline, achieves the best accuracy-efficiency trade-off, with sub-millisecond inference on H7 and 5–7,ms on L4/F4. Model 1, which employs a simpler separable-convolution architecture, is a valid alternative with slightly lower F1. Fixed-threshold detectors, while extremely cheap, are sensitive to inter-subject variation and under-perform on certain positions relative to CNN models.

References

- Adams, J. L.; Dinesh, K.; Snyder, C. W.; Xiong, M.; Tarolli, C. G.; Sharma, S.; Dorsey, E. R.; and Sharma, G. 2020. PD-BioStampRC21: Parkinson's Disease Accelerometry Dataset from Five Wearable Sensor Study.
- Bikias, T.; Iakovakis, D.; Hadjidimitriou, S.; Charisis, V.; and Hadjileontiadis, L. J. 2021. DeepFoG: an IMU-based detection of freezing of gait episodes in Parkinson's disease patients via deep learning. *Frontiers in Robotics and AI*, 8: 537384.
- Borzì, L.; Sigcha, L.; and Olmo, G. 2023. Context Recognition Algorithms for Energy-Efficient Freezing-of-Gait Detection in Parkinson's Disease. *Sensors*, 23(9).
- Chaudhuri, K. R.; Azulay, J.-P.; Odin, P.; Lindvall, S.; Domingos, J.; Alobaidi, A.; Kandukuri, P. L.; Chaudhari, V. S.; Parra, J. C.; Yamazaki, T.; et al. 2024. Economic Burden of Parkinson's Disease: A Multinational, Real-World, Cost-of-Illness Study. *Drugs-Real World Outcomes*, 11(1): 1–11.
- Corrà, M. F.; Atrsaei, A.; Sardoreira, A.; Hansen, C.; Aminian, K.; Correia, M.; Vila-Chã, N.; Maetzler, W.; and Maia, L. 2021. Comparison of laboratory and daily-life gait speed assessment during ON and OFF states in Parkinson's disease. *Sensors*, 21(12): 3974.
- Cui, Y.; Jia, M.; Lin, T.-Y.; Song, Y.; and Belongie, S. 2019. Class-Balanced Loss Based on Effective Number of Samples. In *CVPR*.
- Dini, P.; Diana, L.; Elhanashi, A.; and Saponara, S. 2024. Overview of AI-models and tools in embedded IIoT applications. *Electronics*, 13(12): 2322.
- Fathalla, A.; Li, K.; Salah, A.; and Mohamed, M. F. 2022. An LSTM-based distributed scheme for data transmission reduction of IoT systems. *Neurocomputing*, 485: 166–180.
- He, K.; Zhang, X.; Ren, S.; and Sun, J. 2016. Deep Residual Learning for Image Recognition. In *CVPR*.
- Kianoush, S.; Savazzi, S.; Rampa, V.; Costa, L.; and Tolochenko, D. 2023. A random forest approach to body motion detection: Multisensory fusion and edge processing. *IEEE Sensors Journal*, 23(4): 3801–3814.
- Li, X.; Ghafouri, S.; Ji, B.; Vandierendonck, H.; John, D.; and Nikolopoulos, D. S. 2025. QPART: Adaptive Model Quantization and Dynamic Workload Balancing for Accuracy-aware Edge Inference. *arXiv* preprint *arXiv*:2506.23934.
- Loshchilov, I.; and Hutter, F. 2019. Decoupled Weight Decay Regularization. In *ICLR*.
- Mancini, M.; Shah, V. V.; Stuart, S.; Curtze, C.; Horak, F. B.; Safarpour, D.; and Nutt, J. G. 2021. Measuring freezing of gait during daily-life: an open-source, wearable sensors approach. *Journal of neuroengineering and rehabilitation*, 18: 1–13.
- Muthukrishnan, N.; Abbas, J. J.; and Krishnamurthi, N. 2020. A wearable sensor system to measure step-based gait parameters for parkinson's disease rehabilitation. *Sensors*, 20(22): 6417.

- Sandler, M.; Howard, A.; Zhu, M.; Zhmoginov, A.; and Chen, L.-C. 2018. MobileNetV2: Inverted Residuals and Linear Bottlenecks. In *CVPR*.
- Sigcha, L.; Costa, N.; Pavón, I.; Costa, S.; Arezes, P.; López, J. M.; and De Arcas, G. 2020. Deep learning approaches for detecting freezing of gait in Parkinson's disease patients through on-body acceleration sensors. *Sensors*, 20(7): 1895. Simonyan, K.; and Zisserman, A. 2014. Very Deep Convolutional Networks for Large-Scale Image Recognition. *arXiv:1409.1556*.
- Soumma, S. B.; Mangipudi, K.; Peterson, D.; Mehta, S.; and Ghasemzadeh, H. 2024. Self-Supervised Learning and Opportunistic Inference for Continuous Monitoring of Freezing of Gait in Parkinson's Disease. *arXiv preprint arXiv:2410.21326*.